

Certificate of Analysis

Product Name: **ATP Controls (Low, High, X-High)**

Lot Number: **ATPct050220-L**
ATPct050220-H
ATPct050220-X

Storage Conditions: **-20°C**

Expiration Date: **02/05/21**


FOR RESEARCH USE ONLY

CAUTION: Not intended for human or animal therapeutic use.

Quality Control

Test	Specification	Results
Range	Low—0.043-0.057uM High—0.60-0.80uM X-High---1.49-2.01uM	Pass Pass Pass

Manufactured and tested at Preferred Cell Systems™



Paul Bork, QA Manager
Date: *3/12/2020*